

**State of New York  
Bureau of Weights and Measures  
10B Airline Drive  
Albany, New York 12235  
518-457-3146**

Test # 22117

White - Department  
Yellow - Owner  
Pink - Municipal Office

**Large Capacity Scale Report**

| Jurisdiction<br><u>MADISON COUNTY</u>          |                        |                                  |                                  | Date<br><u>7-12-2022</u>               |                                      |  |                                  |                                     |                        |
|--|------------------------|----------------------------------|----------------------------------|--|--------------------------------------|--|----------------------------------|-------------------------------------|------------------------|
| Establishment Name<br><u>NY HAY SALES</u>      |                        |                                  |                                  | Phone<br><u>315-794-3909</u>           |                                      | Business Type<br><u>FARM</u>                 |                                  |                                     |                        |
| Address<br><u>1 MORSE RD</u>                   |                        |                                  |                                  | Scale Cap x d<br><u>200000 x 20 LB</u> |                                      | CLC/Sect Cap<br><u>70000 LB</u>              |                                  |                                     |                        |
| City, State, Zip<br><u>EARLVILLE, NY 13332</u> |                        |                                  |                                  | WM Lic. (Y/N/NA)<br><u>N</u>           |                                      |  |                                  |                                     |                        |
| Indicator Type<br><u>DIGITAL</u>               |                        | Mfr.<br><u>B-TEK</u>             |                                  | Model<br><u>T403S</u>                  |                                      | S/N<br><u>202450312</u>                      |                                  |                                     |                        |
| Platform Type<br><u>LOAD CELL</u>              |                        | Mfr.<br><u>B-TEK SCALES, LLC</u> |                                  | Model<br><u>BT-7011-200-FESD-SL</u>    |                                      | S/N<br><u>75-08218790</u>                    |                                  |                                     |                        |
| Deck Type/Length<br><u>STEEL / 70 FT</u>       |                        |                                  | # Sections<br><u>4</u>           |  | Approaches<br><u>GRAVEL/CONCRETE</u> |  | Overall Condition<br><u>GOOD</u> |                                     |                        |
| 1 Test Type                                    | 2 Strain Load Reading  | 3 Sub Reading                    | 4 Error @ Sub Load From Col 9    | 5 Sub Load Col(3 -- 4)                 | 6 Known Weight Added                 | 7 Expected Reading Col (2+5+6)               | 8 Indicator Reading              | 9 Error Col (8 - 7)                 | Section Tests          |
| <u>ILZ</u>                                     |                        |                                  |                                  |  | <u>3000</u>                          | <u>3000</u>                                  | <u>3000</u>                      | <u>0</u>                            | Test Load <u>24000</u> |
|  |                        |                                  |                                  |  | <u>3000</u>                          | <u>6000</u>                                  | <u>6000</u>                      | <u>0</u>                            | Reading / Error        |
|  |                        |                                  |                                  |  | <u>3000</u>                          | <u>9000</u>                                  | <u>9000</u>                      | <u>0</u>                            | 1. <u>24000 / 0</u>    |
|  |                        |                                  |                                  |  | <u>3000</u>                          | <u>12000</u>                                 | <u>12000</u>                     | <u>0</u>                            | 2. <u>24000 / 0</u>    |
|  |                        |                                  |                                  |  | <u>3000</u>                          | <u>15000</u>                                 | <u>15000</u>                     | <u>0</u>                            | 3. <u>24000 / 0</u>    |
|  |                        |                                  |                                  |  | <u>3000</u>                          | <u>18000</u>                                 | <u>18000</u>                     | <u>0</u>                            | 4. <u>24000 / 0</u>    |
|  |                        |                                  |                                  |  | <u>3000</u>                          | <u>21000</u>                                 | <u>21000</u>                     | <u>0</u>                            | 5. <u>/</u>            |
|  |                        |                                  |                                  |  | <u>3000</u>                          | <u>24000</u>                                 | <u>24000</u>                     | <u>0</u>                            | 6. <u>/</u>            |
|  |                        |                                  |                                  |  | <u>0</u>                             | <u>0</u>                                     | <u>0</u>                         | <u>0</u>                            | 7. <u>/</u>            |
| <u>SUB1</u>                                    | <u>EMPTY WT TRUCK</u>  | <u>20660</u>                     | <u>0</u>                         | <u>20660</u>                           | <u>20000</u>                         | <u>50660</u>                                 | <u>50640</u>                     | <u>-20</u>                          | Test Load              |
|  |                        |                                  |                                  |  | <u>0</u>                             | <u>0</u>                                     | <u>0</u>                         | <u>0</u>                            | Reading / Error        |
| <u>SUB2</u>                                    | <u>TRACTOR W/ BALE</u> | <u>51640</u>                     | <u>-20</u>                       | <u>51660</u>                           | <u>50660</u>                         | <u>102320</u>                                | <u>102300</u>                    | <u>-20</u>                          | 1. <u>/</u>            |
| <u>DL</u>                                      |                        |                                  |                                  |  | <u>51660</u>                         | <u>51660</u>                                 | <u>51640</u>                     | <u>-20</u>                          | 2. <u>/</u>            |
|  |                        |                                  |                                  |  | <u>0</u>                             | <u>0</u>                                     | <u>0</u>                         | <u>0</u>                            | 3. <u>/</u>            |
|  |                        |                                  |                                  |  |                                      |  |                                  |                                     | 4. <u>/</u>            |
|  |                        |                                  |                                  |  |                                      |  |                                  |                                     | 5. <u>/</u>            |
|  |                        |                                  |                                  |  |                                      |  |                                  |                                     | 6. <u>/</u>            |
|  |                        |                                  |                                  |  |                                      |  |                                  |                                     | 7. <u>/</u>            |
| Sensitivity Load @ 0                           |                        |                                  | Sensitivity Load @ Max Test Load |  |                                      | Weights furnished by<br><u>FREEMAN SCALE</u> |                                  | last certified<br><u>SEPT. 2021</u> |                        |

Test Type: **ILZ** - Increasing Load @ Zero **SUB#** - Substitution **STR** - Strain Load **DL** - Decreasing Load

|  |  |  |  |   |  |
|--|--|--|--|---|--|
| <input type="checkbox"/> Normal Test<br><input type="checkbox"/> Retest (following rejection/repair)<br><input checked="" type="checkbox"/> Retest (following adjustment @ site) |  | <b>Tolerance Applied</b><br><input type="checkbox"/> Maintenance<br><input checked="" type="checkbox"/> Acceptance |  | <b>Official Action</b><br><input checked="" type="checkbox"/> Approved/Sealed<br><input type="checkbox"/> Rejected/Ordered Repaired |  |
| Remarks<br><u>CONFIG 3</u><br><u>CAL 7</u><br>   |  |  |  |   |  |
| Acknowledged<br>   |  | Inspected by<br><u>Marc O'Hara</u>   |  | Inspector Number<br><u>196</u>  |  |